



Textilní zkušební ústav, s.p.

Cejl 480/12, 602 00 Brno, Czech Republic
(Textile Testing Institute)


TESTING LABORATORY NO. 1001

TEST REPORT

AZL 20/ 0555-02

CUSTOMER: LP MAN, s.r.o.
K Zornici 1654/18
957 01 Bánovce nad Bebravou
Slovak Republic

SAMPLE: Aqueous gel ANSILVER® - T containing AgCl
(according to the customer order) Colour: white



SUBJECT OF ASSESSMENT: Determination of antibacterial activity - Agar diffusion plate test

CONDITIONS OF APPLICATION OF THE TEST REPORT:

Test Report contains result of the tests related to the submitted sample only. Sampling has been done by customer. The Report may not be reproduced in the way other than as a complete set. Reproduction of certain parts of the Report is subject to approval of the test laboratory, which has issued it. All information about subcontracted tests results or unaccredited test methods is presented in text part of the test report. This Report is a literal translation of the Czech version.

PREPARED BY:
CHECKED BY:
NUMBER OF PAGES:

P. Jarmiřová
J. Kneysová
3

DATE OF ACCEPTANCE:
6.5.2020

DATE OF EXAMINATION:
11.5. – 12.5.2020

DATE OF ISSUE:
15.5.2020



+420 543 426 720
+420 543 426 742
<http://www.tzu.cz>
azl@tzu.cz



Textilní zkušební ústav, s.p.

AZL 20/ 0555-02

page 3

TEST RESULTS:

Aqueous gel ANSILVER® - T containing AgCl Colour: white					
Characteristics	Test method	Measuring unit	Growth		Assessment
Determination of antibacterial activity - Agar diffusion plate test	EN ISO 20645	Growth	<i>K. pneumoniae</i>	none (IZ < 1 mm)	good effect
			<i>S. aureus</i>	none (IZ < 1 mm)	good effect

IZ ... inhibition zone

Approved:

Petr Nasadil
Head of Testing Laboratory

